



1949

1949

VACUUM-GAUGE TUBE

HARD-GLASS BULB, IONIZATION TYPE

DATA**General:**

Filament, Tungsten:*

Voltage (Approx.) 5 ac or dc volts

Current (Approx.) 3.5 amp

Maximum Tube Length (Including tubulation) 11-1/2"

Maximum Tube Radius 2-3/16"

Maximum Bulb Length 5-1/8"

Maximum Bulb Diameter 2-1/16"

Bulb T-16

Tubulation 1/2" Diameter Hard Glass,
Corning Code 772 NonexOperating Position Vertical with tubulation up or
down; Horizontal, with stem
press in vertical plane


Terminal Arrangement See Outline Drawing

Terminal Lead Connections:

Lead 1 - Common

Lead to
Filaments

Lead 2 - Filament

Lead 3 - Filament
(Spare)

Lead 4 - Grid

Top Lead - Plate

Maximum Ratings, Absolute Values:

FILAMENT VOLTAGE 6.5 max. volts

DC PLATE VOLTAGE DURING OPERATION . . -100 max. volts

DC GRID VOLTAGE DURING OPERATION . . +200 max. volts

VOLTAGE ON GRID & PLATE TIED TOGETHER
DURING DEGASSING (DC OR PEAK AC) . . 650 max. voltsGRID & PLATE DISSIPATION (TOTAL)
DURING DEGASSING . . 150 max. watts

AMBIENT TEMPERATURE DURING OPERATION. 100 max. °C

GAS PRESSURE 0.001 max. mm of Hg

Typical Degassing Conditions:

Grid Connected to Plate

Filament Voltage (AC or DC) . . . 6 6 volts

Grid & Plate Voltage 350 rms 500 dc volts

Grid & Plate
Current (Average) 100 150 ma**Typical Operation:**

DC Plate Voltage -22.5 -22.5 -22.5 volts

* The 1949 contains two filaments, one of which is a spare. Values shown are for either filament operated alone. The filament voltage should be kept as low as possible during degassing because use of a low filament voltage materially increases filament life.

← Indicates a change

MARCH 1, 1954

TUBE DEPARTMENT

DATA

RADIO CORPORATION OF AMERICA, HARRISON, NEW JERSEY

1949



1949


VACUUM-GAUGE TUBE

DC Grid Voltage	+80	+110	+160	volts
Grid Current	10	10	10	ma
Sensitivity	80	110	140	μ a/micron [▲]

Calibration:

See curve on following sheet.

▲ 1 micron = 0.001 mm of mercury.



92CS-6817

MARCH 1, 1954

TUBE DEPARTMENT

RADIO CORPORATION OF AMERICA, HARRISON, NEW JERSEY


DATA



1949

1949

CALIBRATION CURVES



MAR. 11, 1947

TUBE DEPARTMENT
RADIO CORPORATION OF AMERICA, HARRISON, NEW JERSEY

92CM-6851

