



ELECTRONIC ZOOM X-RAY IMAGE INTENSIFIER TUBE TH 9415

HIGH RESOLUTION COMPACT TUBE

BRIGHTNESS GAIN : 6000 X

INPUT FIELDS : 16 cm (6") and 10 cm (4")

TH 9415 is a 16 cm X-ray image intensifier which converts the X-ray pattern into a light image of high brightness and good contrast.

Technological improvements concerning input screen, electronoptics and viewing screen allow for higher resolution than that of older types TH 9411 and TH 9412.

In addition, TH 9415 has an adjustable magnification so to display to the user input fields of 16 cm and 10 cm. The output image is delivered on the output screen with a 14, 2 mm diameter. Magnification can be adjusted from 1/8 to 1/11 by a proper choice of voltages applied to only two focusing electrodes. Output image luminance which is proportional to the energy gain of the intensifier and to the square of the electronic magnification varies from 3000 X to 6000 X.

The "active getter" provides a very high vacuum all along tube's life avoiding the development of an ion spot and alteration of contrast and resolution. The getter needs no extra installation and operates automatically as the tube is energized.

TH 9415 is delivered in a shell which permits mechanical mounting, protects the tube from stray magnetic field and secures a precise positioning of the associated optical system.

Applications

As a result of the high brightness gain it is possible to operate the tube with a very low input dose rate which reduces health hazards. This point is very important in medical radiology since this device enables a significant reduction of X-ray radiation as well on the patient as on the operator.

The high gain allows a good matching with television camera. Such an equipment enables examination of any part of the human body, which gives substantial improvement with respect to conventional X-ray examinations.

Fields of applications in industrial radiology are considerably extended by the use of the tube TH 9415 which allows the use of low power X-rays sources for non destructive testing of very opaque materials.



GENERAL CHARACTERISTICS

Mechanical - (See note 1 and drawing)

The tube is delivered mounted in a metallic shell which permits mechanical fixation : nevertheless it is necessary to maintain the assembly by sustaining the entrance plane. Two types of tapped holes are provided for International or U.S. Standard.



In this shell the tube is fixed so as to facilitate the precise localization of the optical device as regard to the reference plane. This shell protects the tube against stray magnetic field and cannot be considered as an efficient shield against X-radiations. A lead shield of 2 mm thick is provided at the output plane of the shell.

Dimensions	See drawing
Net weight, approximate	4.5 kg
Mounting position	any
Shipment position	tube axis horizontal
Operating and stocking temperature	+50° C
max.	+ 5° C

Optical

Input screen diameter	16 ± 0.5 cm
Output image diameter (O. I.)	14.2 ± 0.5 mm
Input spectral response	X-rays (30 to 250 kVp generator)
Viewing screen	
Type	P 20 ($\lambda = 520$ nm)
Fluorescence and phosphorescence	Yellow - green
Electrostatic focus - Inverted image	

Operating modes :	Normal	Magnified	
Input field diameter	16	10	cm
Magnification	1/11	1/8	
Minimum resolution (note 2)			
- central	20	22	lp/cm
- peripheral	18	20	lp/cm
Minimum contrast (note 3) measured			
with JEDEC penetrometer	3	2.5	%
Minimum conversion factor (note 4)	70	35	Cd/m ² per mR/s
or	350	175	ft. L per R/mn
Minimum luminance gain (note 5)	6000	3000	X
Maximum background luminance (note 6)	0.06	0.06	Cd/m ²
or	0.02	0.02	ft. L.
Maximum distortion (note 7)	20	5	%
Maximum persistence at 10 ms (note 8)	10	10	%



MAXIMUM RATINGS AND TYPICAL OPERATION

Maximum ratings

Photocathode C voltage	0	V
Electrode g1 voltage	0.3	kV
Electrode g2 voltage	2	kV
Electrode g3 voltage	7.5	kV
Anode A voltage	32	kV
Active getter voltage (anode g4)	3.5	kV
(cathode E)	0	V
Photocathode C maximum current	0.5	μ A
Electrode g1 maximum current	7	μ A
Electrode g2 maximum current	1	μ A
Electrode g3 maximum current	1	μ A
Electrode g4 maximum current (after post-gettering operation)	10	μ A
Anode A maximum current	2	μ A

Typical operation (see Starting Procedures)

Operating modes	Normal		Magnified	
Input diameter for 14, 2 mm O. I.	16		10	cm
Photocathode C voltage	0		0	V
Electrode g1 voltage*	100 to 250		100 to 250	V
Electrode g2 voltage*	400 to 1000		0 to 500	V
Electrode g3 voltage*	3.0 to 3.5		6.0 to 6.5	kV
Anode A voltage	28 to 30		28 to 30	kV
Active getter voltage (anode g4)	2.5 to 3		2.5 to 3	kV
(cathode E)	0		0	V
Ripple voltage must not exceed 0, 5 %				

* g1 g2 g3 voltages are defined for an anode voltage of 30 kV.



STARTING PROCEDURE

Important

For tube handling, always prescribe use of security goggles (implosion risk).

Tube mounting

Set tube inside a proper metallic housing to protect operator against X-radiation.

Check for darkness of the tube housing (would the tube be checked without container, it should then be placed in a completely dark room).

The housing in which the tube is contained must be humidity proof and deshydrated. This condition is necessary to eliminate all moisture which could initiate coronas and sparks detrimental to tube operation. Moreover, it prevents dust electrostatic attraction on the viewing face.

Connections

All the connections necessary for tube supply (even for the anode) are made through flexible wires with reference marks.

A 10 Megohms resistance protecting the tube against discharge is provided inside the shell in the anode circuit. Time constant thus produced with tube capacitance contributes to ripple voltage filtering.

In series between each voltage supply and corresponding tube connection, insert in the same way a few Megohms protective resistance.

Focusing adjustments

Voltages supplies can be applied to the electrodes in a short time but with a slope not exceeding 5 kV/ms.

Let tube at rest with voltages applied for potential stabilization before adjusting it. (10 s minimum).

Set a 2 mm spaced metallic wire mesh of 5/10 mm diameter or equivalent (stainless steel, copper) in front of the tube and apply X-rays beam.

For each selected operating mode, electrode g3 voltage must be predetermined, the value of which should be within the limits specified in typical operation. For example $g3 = 3.2$ kV for normal mode and $g3 = 6.2$ kV for magnified mode.

Electrode g3 voltage being selected, adjust electrode g1 and g2 voltages within the specified range. The focusing adjustment shall be made as follows :

- Adjust g1 voltage in order to obtain an image as homogeneous as possible in luminance.
- Adjust g2 voltage in order to obtain the optimum resolution.
- Optimize g1 voltage if necessary.

Focusing adjustments are made respectively for "normal mode" and "magnified mode". It is to notice that a single g1 voltage value can be selected for both modes. This property is valid when photocathode current is used to monitor automatic brightness control device.

Gettering operation

In order to assure a high reliability in operation and to maintain optimum performances of the tube, Instructions for Gettering operation should be strictly applied as defined in separate Data TEV 3006.

The purpose of this operation is to pump residual gas in the tube resulting in an ion spot which lowers the image contrast and resolution.



NOTES

- (1) - The mounting of the tube in its shell (see page 2 and drawing) secures the positioning of the viewing screen at a determined optical distance from the mechanical plane of reference on which may be fixed the optical system. This distance is $20 \text{ mm} \pm 0.25$. This shell assures a parallelism of the two planes with a precision higher than $1/800$ radian. It assures a center of image within 0.5 mm from perpendicular axis of the reference plane defined by an aperture of $120.10 \pm 0.05 \text{ mm}$ diameter.
- (2) - The resolution (as referred to input screen) is measured by using square lead pattern consisting in alternate black and white lines of equal width. Any two adjacent lines are designated as a line pair. The impinging X-radiation is produced by a generator operating at 65 kV and with a 2.5 mm aluminum filter. The resolution decreases gradually from the center to the edge of image. The central zone is within a circle having a diameter equal to 70% of the input diameter. Peripheral zone is within circles 70 and 90% of the input diameter.
- (3) - The contrast is defined as the differential thickness which can be detected when using a JEDEC penetrometer. This penetrometer consists in an aluminum disc of 20 mm thickness presenting holes of 6 mm diameter. The depth of those holes vary from 1.5 to 7% of the thickness of the disc. The differential thickness (expressed in $\%$) of the hole having the minimum depth which can be detected defines the minimum contrast. X-ray conditions : 80 kVp - $\text{HVL } 7 \pm 0.2 \text{ mm Al}$ - input dose rate 20 mR/mn .
- (4) - The conversion factor is the value of the viewing screen luminance corresponding to a determined X-ray dose rate.
X-ray conditions : 80 kVp - 20 mm Al filter - $\text{HVL } 7 \pm 0.2 \text{ mm Al}$.
The luminance is measured by a photometer which matches the human vision.
The conversion factor is defined as :

$$\text{C. F.} = \frac{\text{Luminance}}{\text{dose rate}} = \frac{\text{Candela/square meter}}{\text{milli Roentgen/second}} \quad \text{or} \quad \frac{\text{foot Lambert}}{\text{Roentgen/minute}}$$

- (5) - The luminance gain is the ratio of luminance of the image intensifier to the luminance of a Massiot - Fluor - Sirius fluoroscopic screen having a luminance of 0.012 cd/m^2 per mR/s . Both are irradiated in the same conditions : 80 kVp - 20 mm Al filter - $\text{HVL } 7 \pm 0.2 \text{ mm Al}$. The luminance is measured by a photometer which matches the human vision.
- (6) - The background luminance is the luminance of output screen when normal operating voltages are applied to the tube and X-rays are off at normal ambient temperature.
- (7) - The distortion is measured by putting an object of 1 cm length at the center and then at the edge of the field. The dimensions of the images measured on the viewing screen are respectively l_1 and l_2 . The distortion is given by :



$$D = \frac{l_2 - l_1}{l_1}$$

- (8) - The image persistence is the residual luminance measured at a determined time after removal of X-ray radiations.

OUTLINE DRAWING

ELECTRICAL CONNECTIONS		
Cathode	C	Black
Electrode	g1	Black
Electrode	g2	Black
Electrode	g3	Black
Getter (+)	g4	Black
Getter (-)	E	Blue
Earth	G	Black
Anode	A	White

Cable length 450 about



Dimensions in mm.

