ELECTRONIC TUBE REFERENCE

5519

Specialized - X-Ray Tube electronic tube

CONSOLIDATED TECHNICAL DATA
1 SECTION
CONSTRUCTION
XRAY

Oil-cooled Westinghouse diagnostic X-ray tube with dual 2.1 and 4.2 mm focal spots for high-voltage fluoroscopic and radiographic service.

General characteristics

BASE DESIGNATION	_	**BASE PINS**	_
NOMINAL HEATER	_ / _	**PARALLEL HEATER**	_ / _
HEATING	Direct	**SERIES OPERATION**	Not specified
CATHODE ARRANGEMENT	Separate cathodes	**SECTION CONSTRUCTION**	Similar

Section ratings

SECTION1 XRay

MAXIMUM PLATE VOLTAGE	110000 V

Published ratings and characteristics promoted where clearly applicable; additional source values retained as section parameters.

Westinghouse X-ray tube with 2.1 and 4.2 mm double focus focal spot specification for diagnostic/fluoroscopic high-voltage service.

SOURCE OPERATING CONDITIONS

1947 Westinghouse

SOURCE PDF

Published test conditions and operating points. Values on this page are preserved as reported by this source.

1947 WESTINGHOUSE

Section 1 · XRay

CHARACTERISTICS		
PARAMETER	PUBLISHED VALUE	CONDITION / NOTE
Filament voltage range	**3.5–10 V**	Published
Filament current range	**3.5 A–5.5 A**	Published
Cooling	**Oil**	Published
Focal spot size	**2.1 and 4.2 mm double focus**	Projected size

OTHER		
PARAMETER	PUBLISHED VALUE	CONDITION / NOTE
Heat capacity	**100000 heat unit**	Published
Continuous rating	**30000 heat unit/min**	Published
Peak plate voltage	**110 kV**	Highest published full-wave/self-rectified value

TUBE TINKER

An independent reference for electronic tubes, vacuum tubes, and their source documents.

Catalog Tube types Full books Blog Feedback & corrections

© 2026 Tube Tinker